

Effect of Voxel Size on Finite-Element Analysis of Micro-CT Derived Bone Sample

Upadhyaya K, McErlain D, Sandino C, Boyd S

Abstract

Introduction: Bone strength is dependent on the structural parameters of the trabecular micro-architecture¹. A method to estimate bone strength is finite element (FE) analysis of the bone micro-architecture². Quantification of structural parameters³ and FE analysis results are dependent on the image resolution². This study used micro-computed tomography (micro-CT) to investigate how voxel size affects the accuracy of trabecular bone measurements, particularly regarding how it relates to FE modeling prediction of bone strength.

Methods: Cadaveric bovine cubic bones were imaged at an isotropic voxel size of $20\mu\text{m}$ using a micro-CT scanner (Micro-CT35). Images were segmented using a threshold based technique and re-scaled to voxel sizes 2-4 times larger ($40\mu\text{m}$ - $80\mu\text{m}$) than the original images. Three-dimensional analyses of trabecular bone properties were quantified within the images of the bone cubes. Image voxels were converted to hexahedral elements for FE analysis. Uniaxial 1% compression test was performed on all data (FAIM 5.4). Nodes on the bottom surface were fixed while the top surface was subjected to compression. No constraints were applied to the x and y directions.

Results: Trabecular number (TbN) measurements increased linearly with increasing resolution. There was a 22.11% difference between trabecular number values at $20\mu\text{m}$ versus $80\mu\text{m}$. All other structural parameters were not statistically significant between different image resolutions ($p > 0.05$). For FE analysis, there was a 3.05% percent difference for mean von-Mises Stress at $20\mu\text{m}$ versus $80\mu\text{m}$. Total reaction force between $20\mu\text{m}$ and $80\mu\text{m}$ differed by 0.484%. Maximum von-Mises stress was statistically significantly different between $20\mu\text{m}$ and $80\mu\text{m}$.

Conclusion: All structural parameters except TbN measured at $20\mu\text{m}$ are comparable to $80\mu\text{m}$. Similarly, bone strength estimates through FE analysis at $20\mu\text{m}$ are comparable to $80\mu\text{m}$. It is unlikely that TbN influenced the bone strength estimates. These results will allow for non-invasive estimate of bone strength with advanced clinical CT scanners.

References

1. Dalle Carbonare L, Giannini S (2004) Bone microarchitecture as an important determinant of bone strength. *J Endocrinol Invest* 27:99-105.
2. Bevill G, Keaveny TM (2009) Trabecular bone strength predictions using finite element analysis of micro-scale images at limited spatial resolution. *Bone* 44(4):579-84.
3. Tjong W, Kazakia GJ, Burghardt AJ, Majumdar S. (2012) The effect of voxel size on high-resolution peripheral computed tomography measurements of trabecular and cortical bone microstructure. *Med Phys*. 39(4):1893-903.